

# Calibration Certificate Test Record

CUSTOMER	MACHINE DETAILS	CERTIFICATE	TEST WEIGHTS
Bolton Recycling & Waste Management Bramford Road Great Blakenham Ipswich Suffolk IP6 0SL <u>CONTACT</u>  Jason Smith	<u>MAKE &amp; MODEL</u>  Bilanciai D800 Weighbridge <u>SERIAL NO.</u> T113104431 <u>CAPACITY</u> 50000kg <u>DIVISION</u> 20kg <u>LOCATION</u> In  <u>I.D.</u> <u>TOLERANCE</u> +/- 3 divisions	<u>CERT NO.</u>  581435H <u>TESTED BY</u> Gary Foord <u>DATE TESTED</u> July 2024 <u>NEXT TEST DUE</u> July 2025	<u>SET NO. / CERT NO.</u>  Test Unit - EWT: TR12274/TP0057

### AS FOUND

LOAD POINT	READING	ERROR + / -
0kg	0kg	0kg
2000kg	2040kg	+ 40kg
7000kg	7080kg	+ 80kg
10000kg	10160kg	+ 160kg
20000kg	20340kg	+ 340kg
30000kg	30500kg	+ 500kg
40000kg	40640kg	+ 640kg
50000kg	50800kg	+ 800kg

### DEFINITIVE

LOAD POINT	READING	ERROR + / -
0kg	0kg	0kg
2000kg	2000kg	0kg
7000kg	7000kg	0kg
10000kg	10000kg	0kg
20000kg	20000kg	0kg
30000kg	30000kg	0kg
40000kg	40000kg	0kg
50000kg	50000kg	0kg

### CORNERING AS FOUND

CORNER No.	LOAD POINT	READING	ERROR + / -
1	7000kg	7100kg	+ 100kg
2	7000kg	7100kg	+ 100kg
3	7000kg	7100kg	+ 100kg
4	7000kg	7100kg	+ 100kg
5	7000kg	7100kg	+ 100kg
6	7000kg	7100kg	+ 100kg

### CORNERING DEFINITIVE

CORNER No.	LOAD POINT	READING	ERROR + / -
1	7000kg	7000kg	0kg
2	7000kg	7000kg	0kg
3	7000kg	7000kg	0kg
4	7000kg	7000kg	0kg
5	7000kg	7000kg	0kg
6	7000kg	7000kg	0kg

COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

**Name.** Peter Ord      **Position.** Service Manager

**Signature.**

