

Calibration Certificate Test Record

CUSTOMER	MACHINE DETAILS	CERTIFICATE	TEST WEIGHTS
Bolton Recycling & Waste Management Bramford Road Great Blakenham Ipswich Suffolk IP6 0SL <u>CONTACT</u> Jason Smith	<u>MAKE & MODEL</u> Bilanciai D800 Weighbridge <u>SERIAL NO.</u> T104819451 <u>CAPACITY</u> 50000kg <u>DIVISION</u> 20kg <u>LOCATION</u> Out <u>I.D.</u> <u>TOLERANCE</u> +/- 3 divisions	<u>CERT NO.</u> 581434H <u>TESTED BY</u> Gary Foord <u>DATE TESTED</u> July 2024 <u>NEXT TEST DUE</u> July 2025	<u>SET NO. / CERT NO.</u> Test Unit - EWT: TR12274/TP0057

AS FOUND

LOAD POINT	READING	ERROR + / -
0kg	0kg	0kg
2000kg	2000kg	0kg
7000kg	7000kg	0kg
10000kg	10000kg	0kg
20000kg	19980kg	- 20kg
30000kg	29960kg	- 40kg
40000kg	39900kg	- 100kg
50000kg	49860kg	- 140kg

DEFINITIVE

LOAD POINT	READING	ERROR + / -
0kg	0kg	0kg
2000kg	2000kg	0kg
7000kg	7000kg	0kg
10000kg	10000kg	0kg
20000kg	19980kg	- 20kg
30000kg	29960kg	- 40kg
40000kg	39900kg	- 100kg
50000kg	49860kg	- 140kg

CORNERING AS FOUND

CORNER No.	LOAD POINT	READING	ERROR + / -
1	7000kg	6980kg	- 20kg
2	7000kg	6940kg	- 60kg
3	7000kg	6980kg	- 20kg
4	7000kg	7000kg	0kg
5	7000kg	6940kg	- 60kg
6	7000kg	6960kg	- 40kg

CORNERING DEFINITIVE

CORNER No.	LOAD POINT	READING	ERROR + / -
1	7000kg	6980kg	- 20kg
2	7000kg	6940kg	- 60kg
3	7000kg	6980kg	- 20kg
4	7000kg	7000kg	0kg
5	7000kg	6940kg	- 60kg
6	7000kg	6960kg	- 40kg

COMMENTS OR OBSERVATIONS
Recommend cleaning out.

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord **Position.** Service Manager **Signature.** 