

# Calibration Certificate Test Record

| CUSTOMER                                                                                                                                          | MACHINE DETAILS                                                                                                                                                                                                              | CERTIFICATE                                                                                                                                 | TEST WEIGHTS                                                             |
|---------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------|
| Bolton Recycling & Waste Management<br>Bramford Road<br><br>Great Blakenham<br>Ipswich<br>Suffolk<br>IP6 0SL<br><u>CONTACT</u><br><br>Jason Smith | <u>MAKE &amp; MODEL</u><br><br>Bilanciai D800 Weighbridge<br><br><u>SERIAL NO.</u> T113104431<br><u>CAPACITY</u> 50000kg<br><u>DIVISION</u> 20kg<br><u>LOCATION</u> In<br><br><u>I.D</u><br><u>TOLERANCE</u> +/- 3 divisions | <u>CERT NO.</u><br><br>584568H<br><br><u>TESTED BY</u><br>Peter Ord<br><u>DATE TESTED</u><br>July 2025<br><u>NEXT TEST DUE</u><br>July 2026 | <u>SET NO. / CERT NO.</u><br><br>Test Unit: EWT Weights<br>TR12695/98182 |

## AS FOUND

| LOAD POINT | READING | ERROR + / - |
|------------|---------|-------------|
| 0kg        | 0kg     | 0kg         |
| 2000kg     | 2000kg  | 0kg         |
| 7000kg     | 7000kg  | 0kg         |
| 10000kg    | 10000kg | 0kg         |
| 20000kg    | 20000kg | 0kg         |
| 30000kg    | 30000kg | 0kg         |
| 40000kg    | 40000kg | 0kg         |
| 50000kg    | 50000kg | 0kg         |
|            |         |             |
|            |         |             |
|            |         |             |
|            |         |             |

## DEFINITIVE

| LOAD POINT | READING | ERROR + / - |
|------------|---------|-------------|
| 0kg        | 0kg     | 0kg         |
| 2000kg     | 2000kg  | 0kg         |
| 7000kg     | 7000kg  | 0kg         |
| 10000kg    | 10000kg | 0kg         |
| 20000kg    | 20000kg | 0kg         |
| 30000kg    | 30000kg | 0kg         |
| 40000kg    | 40000kg | 0kg         |
| 50000kg    | 50000kg | 0kg         |
|            |         |             |
|            |         |             |
|            |         |             |
|            |         |             |

## CORNERING AS FOUND

| CORNER No. | LOAD POINT | READING | ERROR + / - |
|------------|------------|---------|-------------|
| 1          | 7000kg     | 7000kg  | 0kg         |
| 2          | 7000kg     | 7000kg  | 0kg         |
| 3          | 7000kg     | 7000kg  | 0kg         |
| 4          | 7000kg     | 7000kg  | 0kg         |
| 5          | 7000kg     | 7000kg  | 0kg         |
| 6          | 7000kg     | 7000kg  | 0kg         |
|            |            |         |             |
|            |            |         |             |

## CORNERING DEFINITIVE

| CORNER No. | LOAD POINT | READING | ERROR + / - |
|------------|------------|---------|-------------|
| 1          | 7000kg     | 7000kg  | 0kg         |
| 2          | 7000kg     | 7000kg  | 0kg         |
| 3          | 7000kg     | 7000kg  | 0kg         |
| 4          | 7000kg     | 7000kg  | 0kg         |
| 5          | 7000kg     | 7000kg  | 0kg         |
| 6          | 7000kg     | 7000kg  | 0kg         |
|            |            |         |             |
|            |            |         |             |

COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

**Name.** Peter Ord      **Position.** Service Manager      **Signature.**

