

Calibration Certificate Test Record

CUSTOMER	MACHINE DETAILS	CERTIFICATE	TEST WEIGHTS
Bolton Recycling & Waste Management Bramford Road Great Blakenham Ipswich Suffolk IP6 0SL <u>CONTACT</u> Jason Smith	<u>MAKE & MODEL</u> Bilanciai D800 Weighbridge <u>SERIAL NO.</u> T104819451 <u>CAPACITY</u> 50000kg <u>DIVISION</u> 20kg <u>LOCATION</u> Out <u>I.D.</u> <u>TOLERANCE</u> +/- 3 divisions	<u>CERT NO.</u> 584567H <u>TESTED BY</u> Peter Ord <u>DATE TESTED</u> July 2025 <u>NEXT TEST DUE</u> July 2026	<u>SET NO. / CERT NO.</u> Test Unit: EWT Weights TR12695/98182

AS FOUND

LOAD POINT	READING	ERROR + / -
0kg	0kg	0kg
2000kg	2000kg	0kg
7000kg	7000kg	0kg
10000kg	9980kg	- 20kg
20000kg	19960kg	- 40kg
30000kg	29920kg	- 80kg
40000kg	39880kg	- 120kg
50000kg	49840kg	- 160kg

DEFINITIVE

LOAD POINT	READING	ERROR + / -
0kg	0kg	0kg
2000kg	2000kg	0kg
7000kg	7000kg	0kg
10000kg	10000kg	0kg
20000kg	20000kg	0kg
30000kg	30000kg	0kg
40000kg	40000kg	0kg
50000kg	50000kg	0kg

CORNERING AS FOUND

CORNER No.	LOAD POINT	READING	ERROR + / -
1	7000kg	7000kg	0kg
2	7000kg	7000kg	0kg
3	7000kg	7000kg	0kg
4	7000kg	7000kg	0kg
5	7000kg	7000kg	0kg
6	7000kg	7000kg	0kg

CORNERING DEFINITIVE

CORNER No.	LOAD POINT	READING	ERROR + / -
1	7000kg	7000kg	0kg
2	7000kg	7000kg	0kg
3	7000kg	7000kg	0kg
4	7000kg	7000kg	0kg
5	7000kg	7000kg	0kg
6	7000kg	7000kg	0kg

COMMENTS OR OBSERVATIONS

Debris found under end of bridge causing As Found errors, cleared and all OK

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord **Position.** Service Manager **Signature.**

