

Calibration Certificate Test Record

CUSTOMER	MACHINE DETAILS	CERTIFICATE	TEST WEIGHTS
Bolton Recycling & Waste Management Bramford Road Great Blakenham Ipswich Suffolk IP6 0SL <u>CONTACT</u> Jason Smith	<u>MAKE & MODEL</u> Bilanciai D800 Weighbridge <u>SERIAL NO.</u> T104819451 <u>CAPACITY</u> 50000kg <u>DIVISION</u> 20kg <u>LOCATION</u> Out <u>I.D</u> <u>TOLERANCE</u> +/- 3 divisions	<u>CERT NO.</u> 587607H <u>TESTED BY</u> George Woodruffe <u>DATE TESTED</u> July 2026 <u>NEXT TEST DUE</u> July 2027	<u>SET NO. / CERT NO.</u> Test Unit: EWT Weights TR13194

AS FOUND

LOAD POINT	READING	ERROR + / -
0kg	0kg	0kg
2000kg	2000kg	0kg
7000kg	7000kg	0kg
10000kg	10000kg	0kg
20000kg	20000kg	0kg
30000kg	29980kg	- 20kg
40000kg	39980kg	- 20kg
50000kg	49980kg	- 20kg

DEFINITIVE

LOAD POINT	READING	ERROR + / -
0kg	0kg	0kg
2000kg	2000kg	0kg
7000kg	7000kg	0kg
10000kg	10000kg	0kg
20000kg	20000kg	0kg
30000kg	29980kg	- 20kg
40000kg	39980kg	- 20kg
50000kg	49980kg	- 20kg

CORNERING AS FOUND

CORNER No.	LOAD POINT	READING	ERROR + / -
1	7000kg	7000kg	0kg
2	7000kg	7000kg	0kg
3	7000kg	7000kg	0kg
4	7000kg	7000kg	0kg
5	7000kg	7000kg	0kg
6	7000kg	7000kg	0kg

CORNERING DEFINITIVE

CORNER No.	LOAD POINT	READING	ERROR + / -
1	7000kg	7000kg	0kg
2	7000kg	7000kg	0kg
3	7000kg	7000kg	0kg
4	7000kg	7000kg	0kg
5	7000kg	7000kg	0kg
6	7000kg	7000kg	0kg

COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord **Position.** Service Manager **Signature.**

